

Biological Verification: A.A. (tester's Initials)								
Test date	ear	Test Frequency (Hz) Threshold Result in dB(HL)						
		500	1000	2000	3000	4000	6000	8000
July 30, 2019	L	5	0	0	5	10	15	15
Baseline	R	0	5	0	0	5	10	15
April 17, 2020	L	5	5	0	5	5	10	15
Regular Test Day	R	5	5	10	5	20*	30*	30*

\* frequencies outside of the accepted range of variation of 10 dB(HL)